

MAHARSHI DAYANAND UNIVERSITY ROHTAK
(Established under Haryana Act No. 25 of 1975)
'A+' Grade University accredited by NAAC

DEPARTMENT OF FORENSIC SCIENCE

No.FS./2022/ 1252-1256
Dated;- 15/12/22

To,

o/c

Dear Sir/Madam

Please quote for article mentioned below in the list attached herewith so as reach undersigned

by 7th January, 2023

THE QUALITY, MAKE, PACKING and other particulars of each item should be stated.

WHILE SUBMITTING THE QUOTATIONS PLEASE NOTE THE POINTS MENTIONED BELOW:-

1. The quotations are to be sent in an envelope marked QUOTATION DUE ON 7 Jan, 2023 and this envelope may be put in another envelope addressed to the undersigned.
2. The rates of S.T./ C.S.T., EXCISE, CUSTOM and other takes chargeable must be specified.
3. The rates of insurance, if any, should be specified. The firm will be required to submit the original Payee Receipt alongwith the bill.
4. Please state the time period within which the items will be supplied.
5. GUARANTEE PERIOD OF THE INSTRUMENTS MUST BE MENTIONED.
6. Discount if any, is to be given with the quotation.

Sr. No.	Name of Item	Other Specifications
1	Muffle Furnace	<ul style="list-style-type: none">• Digital Muffle Furnace of Size 150 x 150 x 300mm• Maximum Temp. Range 1200 degree Celsius• Should come with High temp. gloves, Crucible tong, and zirconium oxide and Alumina tray crucible. Warranty- 1 year extendable up to two years
2	Vertical Deep Freezer (-20 degrees Celsius)	<ul style="list-style-type: none">• Temperature Range (C): $\leq -20^{\circ}\text{C}$, Gross Volume -344 (12.2), Net Volume -310 (11.0) L and High-temperature alarm with External digital display single Solid Door• Should come with Digital Voltage Stabilizer. Warranty- 1 year extendable up to two years
3	Microtome	<ul style="list-style-type: none">• Effortless manual sectioning via a counter-balanced, exceptionally smooth-running handwheel and Type Rotary Microtome• Featured with one razor 120 mm and honing stone and Feed range 1-50 micron in steps of one micron with an automatic feed release device and Section 0.5-60 um• Frequency 50 Hz and Voltage 220 V• Should come with an extra razor blade
4	Homogenizer	<ul style="list-style-type: none">• Heavy cast iron with mild steel chrome plated upright rod with adjustable clamp to hold 1/2 HP motor stainless steel stirring rod mounted with Teflon grinder. Complete with adjustable clamp to hold tissue homogenizing of borosilicate glass, Variable speed regulator. To work on 220V AC• Material Stainless Steel and Shaft Diameter 10 mm.• Should come with additional Impeller Blades and Jar of Size 5 ml, 15 ml, 25 and 100 ml

Yours Sincerely,

15/12/22

Head, Deptt. Of Forensic Science
M.D. University, Rohtak
Department of Forensic Science
Maharshi Dayanand University
Rohtak-124001 (Haryana)