

MAHARSHI DAYANAND UNIVERSITY, ROHTAK

Corrigendum

Tender for PURCHASE OF LABORATORY INSTRUMENTS in M.D.U. Rohtak

Tender Reference Number: DST

Tender ID: 2019_HRY_92907_1

1. Page No. 2 Tender Docs Fee+ E Service Fees may be read as Rs. 2000 + 1000 = 3000/-
2. Page No. 16, Item no. 23 is added which may be read as; it should be supplied with a suitable servo stabilizer.
3. Page No. 17, Item no. 21 is added which may be read as; it should be supplied with a suitable servo stabilizer.
4. Page No. 18, Item no. 25 is added which may be read as; it should be supplied with a suitable servo stabilizer.
5. Page No. 19, Item no. 1 may be read as Max. Speed : 4000rpm or more (with 50 ml rotor) and 17000rpm or more (with 2ml/1.5ml rotor)
6. Page No. 19, Item no. 13 may be read as Built in Condensation drain to eliminate water accumulation and prevent corrosion optional.
7. Page No. 19, Item no. 17 may be read as Fixed angled rotor, 6x50ml and 6x15ml with adaptors, RPM should be 9000 or more.
8. Page No. 19, Item no. 18 may be read as Fixed angled rotor 24 x 2ml Rotor, RPM should be 15000 or more.
9. Page No. 19, Item no. 19 may be read as Compatible Servo Stabilizer.

Anita
23/8/19
Dr. ANITA RANI SANTAL
PI-DST-SERB
Assistant Professor
Department of Microbiology
M. D. University, Rohtak-124001

Praveen
Praveen
23/8/19

Head
Department of Microbiology
M.D. University,
ROHTAK-124001

G. S. Singh
27/08/19
REGISTRAR