

**DEPARTMENT OF MICROBIOLOGY
MAHARSHI DAYANAND UNIVERSITY ROHTAK**

No. Micro/2025/850

Dated: 22/12/2025

Last date of sending Quotation: 05/01/2026

M/s.


Dear Sir/Madam,

Please quote for the article mentioned below so as reach the undersigned by 05/01/2026. The quality, make, packaging and other particulars of each item should be stated.

WHILE SUBMITTING THE QUOTATIONS PLEASE NOTE THE POINTS MENTIONED BELOW:

1. The quotations are to be sent in an envelope marked QUOTATION DUE ON 05/01/2026 and this envelop may be put in another envelope addressed to the undersigned.
2. Quotation are invited from authorized dealers with authorization certificate.
3. The rates of S.G.S T./C. G.S.T., EXCISE, CUSTOM and another taxes chargeable must be specified.
4. Discount if any, is to be given with the quotations.
5. The price quoted will be applicable to the specification mentioned and will remain valid for at least 90 days.
6. The selected vendor will be required to supply and install the items within two weeks from the date of issue of purchase order.
7. No advance payment shall be made; any offer linked with advance payment is likely to be ignored/rejected.
8. Incomplete offers will be summarily rejected. The warranty period will be reckoned from the date of installation and acceptance of the items by MD UNIVERSITY, ROHTAK
9. Quotation sent must include Quotation Ref.NO. in subject line followed by name of vendor j. The acceptance of goods/services is subject to the approval of the Inspection Committee.
10. Dispute if any, will be subject to Rohtak Jurisdiction.

Sr. No.	Name of item
1.	Benchtop Orbital Shaking Incubator (REMI) (Details as per list attached)


22.12.2025
HOD (Microbiology)

Head
Department of Microbiology
M.D. University,
Rohtak-124001

Specifications for Benchtop Orbital Shaking Incubator (REMI)

SALIENT FEATURES:

- Stainless-steel chamber and platform easy to clean
- Transparent plastic door
- Microprocessor controller with large LCD screen for temperature and shaking speed with timing function
- Self-check function easy to identify problems
- Smooth start and stop system prevents liquid spillage
- Auto-controller of fan speed to prevent damage to the samples
- Safety door switch, auto pause operation when it is opened

Technical Data

Shaking Speed Range	20-300rpm
Amplitude	20mm
Temperature Range	4-50°C $\pm 0.5^{\circ}\text{C}$
Display Resolution	0.1°C
Timer Range	1 hour 99h:59m
Platform Size (mm)	450 x 450
Maximum Shaking Capacity	4x2000ml, 8x1000ml, 11x500ml,
(No. of flasks x Volume)*	12x250ml, 28x100ml, 42x50ml
Load in Tray	10 Kg
Power Consumption	1100w
Recommended Voltage Stabilizer	VS 02