

MAHARSHI DAYANAND UNIVERSITY ROHTAK
(Established under Haryana Act No. 25 of 1975)
'A+' Grade University accredited by NAAC
DEPARTMENT OF GENETICS

No.Genet./2022/ 1081-83
Dated:- 20-12-2022

To,

6/C

Dear Sir/Madam

Please quote for article mentioned below in the list attached herewith so as reach undersigned by 5.1.2023.

THE QUALITY, MAKE, PACKING and other particulars of each item should be stated. WHILE SUBMITTING THE QUOTATIONS PLEASE NOTE THE POINTS MENTIONED BELOW:-

1. The quotations are to be sent in an envelope marked QUOTATION DUE ON Jan 5, 2023 and this envelope may be put in another envelope addressed to the undersigned.
2. The rates of S.T./ C.S.T., EXCISE, CUSTOM and other takes chargeable must be specified.
3. The rates of insurance, if any, should be specified. The firm will be required to submit the original Payee Receipt alongwith the bill.
4. Please state the time period within which the items will be supplied.
5. GUARANTEE PERIOD OF THE INSTRUMENTS MUST BE MENTIONED.
6. Discount if any, is to be given with the quotation.

Sr. No.	Name of Item/ Specifications	Make	Model/ Cat no.	Other Specifications
1	<u>Nano-Drop Spectrophotometer</u> Used for quantification of nucleic acid, DNA, RNA & proteins. Technical specifications:- Silicon photodiode detector, Absorbance Range of 0 - 30 Abs with detection limits 0.12 mg/ml to 45 mg/ml (BSA)4 ng/μ, Light Emitting Diodes (LED), sample used (1)with sample volume(1-2μL),Spectral Resolution(<8.0 nm),Volume (Metric,1μL) Wavelength Repeatability with 0.002 *SD of 10 individual measurements at 0.74Abs including all accessories for functional unit. Power Supply: Should include voltage (6V), Wattage with 18w including all accessories for functional unit & online UPS Certification: Unit should be UL/CSA and CE Friendly user interface ,Display all important parameters with LCD Services: Required: Installation & on site validation, Calibration certificates Manuals: Operation, maintenance & part list with detailed Specifications Operational & maintenance Training Warranty of 1 year standard extendable to 2 years extra.			

Yours Sincerely,

Minehal
20.12.22
Head, Deptt. Of Genetics
Department of Genetics
M.D. University, Rohtak
Rohtak - 124001 (Haryana)