MAHARSHI DAYANAND UNIVERSITY, ROHTAK NEXT GENERATION GENOME SEQUENCING LABORATORY

No. MDU/NGSL/2022/172

Dated: 12th October 2022

Quotation Inviting Notice

Quotations are invited from interested distributors/suppliers having GST number for the supply of the Orbital shaker Incubator.

Sr. No.	Name of Item	Required	Specifications/Make/etc.
140.		Qty.	
1	Orbital shaker	One	Table top, Operating temperature: 15 to 50°C, Temperature
	Incubator		accuracy \pm 0.2°C, Temperature uniformity across cabinet \pm 1.0C,
			Speed: 25-300 rpm or higher, speed accuracy: ±1, Orbit: 19-
			25mm, Timer, Flexible orbital shaking platform to accommodate
			flasks of varying capacity (100ml, 250ml, 500ml amd 1000ml),
			Platform should accommodate at least 24 of 100ml flasks (or
			higher) at a time, clamps for 100ml (24no), 250ml (26no), 500ml (8
			no) and 1000ml (4 no), Additional stainless steel rack (01) to
			incubate culture plates, Proper internal lighting system, Supplied
			with appropriate voltage stabilizer, Onsite warranty: Two year

Please quote for the articles mentioned above in the list attached herewith to reach the undersigned by 5.00 pm on 27th October 2022.

The Quality, Make, Packaging, and other particulars of each item should be stated while submitting the quotation. Please note the points mentioned below:

- 1. The quotations are to be sent in the sealed envelope marked "QUOTATION OF ORBITAL SHAKER INCUBATOR DUE ON 27TH OCTOBER 2022" and addressed to the undersigned.
- 2. The rates of GST and other taxes chargeable must be specified.
- 3. The rate of Insurance, if any, should be specified. The firm will be required to submit the original payment receipt along with the bill.
- 4. Please state the time period within which the item will be supplied.
- 5. GUARANTEE PERIOD of the instrument must be mentioned
- 6. Discount if any, is to be given with the quotation.

Project Coordinator

Dr. Nar Singh Chauhan Principal Investigator

Next Generation Genome Sequencing Laboratory Maharshi Dayanand University, Rohtak - 124001