

MAHARSHI DAYANAND UNIVERSITY ROHTAK
(Established under Haryana Act No. 25 of 1975)
'A+' Grade University accredited by NAAC
DEPARTMENT OF GENETICS

No.Genet./2022/ _____
Dated:- _____

To, _____

Dear Sir/Madam


Please quote for article mentioned below in the list attached herewith so as reach undersigned by 24-11-2022

THE QUALITY, MAKE, PACKING and other particulars of each item should be stated.
WHILE SUBMITTING THE QUOTATIONS PLEASE NOTE THE POINTS MENTIONED BELOW:-

1. The quotations are to be sent in an envelope marked QUOTATION DUE ON 24/11/22 and this envelope may be put in another envelope addressed to the undersigned.
2. The rates of S.T./ C.S.T., EXCISE, CUSTOM and other takes chargeable must be specified.
3. The rates of insurance, if any, should be specified. The firm will be required to submit the original Payee Receipt alongwith the bill.
4. Please state the time period within which the items will be supplied.
5. GUARANTEE PERIOD OF THE INSTRUMENTS MUST BE MENTIONED.
6. Discount if any, is to be given with the quotation.

Sr. No.	Name of Item/ Specifications	Make	Model/ Cat no.	Other Specifications
1	<p><u>Inverted Microscope with Digital Camera & Work Station</u> (Camera & Microscope should be from same manufacture) <u>Objectives</u> : Infinity Plan 4x phase , & Infinity Plan LWD 10XPH, 20XPH & LWD 40XPH <u>Condenser</u>: Long Working Distance, N.A. 0.30, WD 70mm with Phase Slider for 10X-40X & Filter Slot <u>Nosepiece</u>: Nosepiece with capacity to hold 5 objectives <u>Eyepiece</u>: WF 10x (F.O.V.22) with rubber eye guard. The eyepiece tube interpupillary distance should be 50-75mm, with Inclination of 30-45° <u>Illuminator</u>: Transmitted High luminescent white LED illuminator (lifetime approx.25,000 hrs.) <u>Mechanical stage</u>: Withholder for two sides stage , petri dish holder & teraski holder with XY control drive <u>Photographic attachment</u> Trinocular model to accommodate image documentation, photo port that accepts various micrographic system. Scientific grade Digital camera with 6 mega pixel CMOS/CCD high resolution, Multiple Fluorescence mode , Annotation & Measurement on live & captured images, Should able to Capture in Images & Video/Time lapse mode, Calibration Bar on live/captured images, Etc . Data Station: Latest Branded Corei3 Desktop with 20" HD LED Monitor, 4GB RAM 1TB Hard Disk, Keyboard & Optical Mouse, USB 3.0/Firewire Port for camera Power supply: Should include 210-240V/50-60 Hz Spares : Lamps : 02 Nos. Services: Required: Installation & on site validation Manuals: Operation, maintenance & part list with detailed Specifications Operational & maintenance Training Warranty of 1 year standard + Optional 2 years extra.</p>			

Yours Sincerely,


Head, Deptt. Of Genetics
Maharshi Dayanand University,
Rohtak - 124001 (Haryana)
03.11.22